

# GLUCOSE (5 MIN.)

**(GOD/POD Method)**

Reagent kit for quantitative estimation of glucose in serum or plasma.

**Product Highlights :**

- Low blank .
- End Point formulation with five minutes incubation.
- Suitable pack size as per customer need.
- Linearity & Accuracy as per international standard.



**BACKGROUND AND SYNOPSIS :**

Conventional methods like Ferricyanide, Nelson-Somogyii, Folin-Wu, O-Tolidine involved lengthy procedures and were cumbersome. Some of these methods involved steps like precipitation, filtration and boiling. These methods have drawbacks, are non-specific, imprecise, insensitive and inaccurate.

**ENZOPAK** Glucose is based on GOD/POD method (Glucose oxidase / Peroxidase) as described by Trinder and is specific and accurate. The method selectively estimates β -D-Glucose in serum or plasma.

**PRINCIPLE :**

Glucose oxidase oxidises the specific substrate, β -D-glucose, to gluconic acid and hydrogen peroxide is generated. Hydrogen-peroxide thus produced is acted upon by peroxidase and oxygen is liberated. The liberated oxygen is transferred to chromogen system consisting of 4-aminoantipyrine and phenolic compound to produce red quinoneimine dye. The intensity of colour is directly proportional to the concentration of glucose and is measured photometrically at 505 nm (500-540 nm or with GREEN filter).



**DIAGNOSTIC SIGNIFICANCE :**

Blood glucose estimations are generally carried out for the diagnosis and follow up of diabetes mellitus.

Glucose circulating in blood plasma is in dynamic equilibrium with absorption of carbohydrates from intestine. High levels of Glucose induces glycogen synthesis and low levels generate carbohydrate, in turn, glucose. Similar pathways are also encountered with Fatty acids and Amino acids (protein). Deficiency of insulin-like activity is a condition resulting in hyperglycemia. Hyperglycemia and Hypoglycemia are also associated with various hormonal disorders e.g. Pituitary, Thyroid etc.

**REAGENT COMPOSITION**

Active Ingredients	Concentration
<b>Reagent-1</b>	
• Buffer	100mmol/L
• GOD	≥ 20000 U/L
• POD	≥ 2500 U/L
• 4 - AAP	0.25mmol/L
pH 7.4± 0.5 at 25°C	

**Reagent-2**

- Phenol 10 mmol/L

**Glucose Standard (100 mg/dl)**

Also contains non-reactive fillers and Stabilizers.

**PRESENTATION:**

Pack Size	No. of Bottles		
	5 x 100 ml	4 x 250 ml	4x500 ml
• 1 Glucose (Enzyme /Chromogen)	5	4	4
• 2 Glucose (Phenol)	1	4	4
• Standard (100 mg/dl)	1	1	2
• Reconstitution Bottle	1	-	-

\* 2 Glucose Phenol (Ready to use) is provided separately.

**PRECAUTION :**

ENZOPAK Glucose is for *IN-VITRO* diagnostic use only.  
**Reagent Contains Sodium Azide. DO NOT INGEST.**

**PREPARATION OF WORKING REAGENT**

**FOR 5 x 100 ml. :**

Transfer the contents (powder mixture of Enzyme/Chromogen) of one bottle of 1 Glucose to the bottle provided for reconstitution. To this add 100 ml. of 2 Glucose Phenol "Ready to Use". Mix well to dissolve and use as given in the procedure. Store at 2-8° C when not in use.

**4 x 250 ml. :**

Transfer the content (powder mixture of Enzyme/Chromogen) of one bottle of 1 Glucose to the bottle of 2 Glucose Phenol "Ready to Use". Mix well to dissolve and use as given the procedure. Store at 2-8°C when not in use.

**4 x 500 ml. :**

Transfer the contents (powder mixture of Enzyme/Chromogen) of 1 pouch/ bottle of 1 Glucose to the bottle of 2 Glucose Phenol "Ready to Use". Mix well to dissolve and use as given the procedure. Store at 2-8°C when not in use.

**REAGENT STORAGE AND STABILITY :**

1 GLUCOSE	2-8°C	Until Expiry
2 GLUCOSE (Phenol)	R.T.	Until Expiry
GLUCOSE STD. (100 mg/dl)	2-8°C	Until Expiry
Working Reagent	2-8°C	90 days
(Protected from light)		

**SPECIMEN COLLECTION :**

Blood sample collected with any one of the anticoagulants like fluoride, oxalate, EDTA, heparin or without any of the anticoagulants can be used. As soon as the sample is collected, separate serum or plasma to prevent glycolysis.

**REACTION PARAMETERS : (END POINT)**

- Type of Reaction : End Point
- Wavelength : 505 nm, (500-540 nm)
- Flowcell Temperature : 37°C
- Incubation : 5 min. at 37°C
- Std. Concentration : 100 mg/dl
- Sample Volume : 10 µl (0.01 ml)
- Working Reagent Volume : 1.0 ml
- Zero setting with : Reagent Blank
- Light Path : 1.0 cm.



# GLUCOSE (5 MIN.)

ENZOPAK

## PROCEDURE:

### For Instruments with 1 ml. cuvette capacity-

PIPETTE INTO TEST TUBES	BLANK	STANDARD	TEST
• WORKING REAGENT (ml)	1.0	1.0	1.0
• STANDARD (ml)	-	0.01	-
• SAMPLE (ml)	-	-	0.01

Mix well and incubate for 5 Min. at 37° C. Mix & read absorbance of standard and test at 505 nm (500-540 nm) or with GREEN filter against reagent blank.

### For instruments with 3 ml. cuvette capacity-

PIPETTE INTO TEST TUBES	BLANK	STANDARD	TEST
• WORKING REAGENT (ml)	1.5	1.5	1.5
• STANDARD (ml)	-	0.02	-
• SAMPLE (ml)	-	-	0.02

Mix well and allow to stand for 5 min. at 37° C. Then add....

• DISTILLED WATER (ml)	1.5	1.5	1.5
------------------------	-----	-----	-----

Mix. & read absorbance of standard and test at 505 nm (500-540 nm or with GREEN filter) against reagent blank.

## STABILITY OF FINAL REACTION MIXTURE :

The colour of reaction mixture is stable for 2 hours at room temperature, when protected from direct light.

## TEST RESULTS :

$$\text{Glucose Concentration (mg/dl)} = \frac{\text{Absorbance of Test}}{\text{Absorbance of Std.}} \times 100$$

## REACTION PARAMETERS (FIX TIME)

- Type of Reaction : Fix Time/Two Point/Initial Rate
- Wavelength : 505 nm (500-540)
- Flowcell Temperature : 37°C
- Delay time : 30 seconds
- Interval time : 60 seconds
- Measuring time : 90 seconds
- No. of Readings : 2
- Standard Sample Volume : 10 µl (0.01 ml)
- Standard Concentration : 100 mg/dl
- Reagent Volume : 1.0 ml
- Zero Setting with : Distilled Water
- Light Path : 1.0 cm.

## PROCEDURE :

PIPETTE INTO TEST TUBES	STANDARD (ST)	TEST (TS)
• WORKING REAGENT(ml)	1.0	1.0
• STANDARD (ml)	0.01	-
• SAMPLE (ml)	-	0.01

Mix and aspirate. Record the absorbance of standard (ST) and Test (TS) at 30 seconds (ST<sub>1</sub>, TS<sub>1</sub>) and again at 90 seconds (ST<sub>2</sub>, TS<sub>2</sub>) at 505 nm, against distilled water.

## TEST RESULTS:

$$\text{Glucose Concentration (mg/dl)} = \frac{(TS_2 - TS_1)}{(ST_2 - ST_1)} \times 100$$

To convert mg/dl to mmol / lit. use the following factor –

$$1 \text{ mmol / lit.} = 18 \text{ mg/dl}$$

$$1 \text{ mg/dl.} = 0.056 \text{ mmol/lit}$$

**LIMITATIONS FOR INTERFERENCE :** As per studies carried out for interference. Following results were obtained.

- No Interference from Hemoglobin upto 187.5 mg/dl.
- No Interference from free Bilirubin upto 25.0 mg/dl.
- No Interference from Lipemic (Measured as Triglycerides) upto 1000 mg/dl.

## NORMAL VALUES :

- Fasting : 70-110 mg/dl (3.90-6.11 mmol/lit)
- Two Hours Post prandial : upto 140 mg/dl (7.78 mmol/lit)

## LINEARITY :

This method is linear upto 500 mg/dl. For sample value above 500 mg/dl, dilute the sample suitably with 0.9 % saline & repeat the assay. Apply correction due to dilution to arrive at a final result.

## REFERENCES :

- TRINDER P., Annual Clinical Biochem 6, 24-25 (1969)
- HENRY. R. J. CANNON D.C., WINKELMAN I. W. Clinical Chemistry, Principles and Techniques 2nd edition. Harper & Row Publiser Inc. , N. Y., P-1288 (1974).
- TIETZ, N. W. (ed.) Fundamentals of Clinical Chemistry, 2nd edition. W. B. Saunders Co., Toronto, (1982).

