

RF LATEX

Rapid latex agglutination slide test for the qualitative and semiquantitative *in-vitro* determination of Rheumatoid factors (RFs) in undiluted serum for the detection of Rheumatoid Arthritis

BACK GROUND & SYNOPSIS:

Many of the major rheumatological disorders are autoimmune diseases; Rheumatoid Arthritis is one of them. An antigenic stimulus (antigen unknown) leads to the appearance of an abnormal IgG that results in the production of rheumatoid factors and eventual development of rheumatoid disease. These factors are present in serum and synovial fluid of the subject. The 'RFs' are also known as anti-immunoglobulins.

PRINCIPLE:

Polystyrene Latex particles are coated with purified human globulin (IgG). When a serum with rheumatoid factors is mixed with latex, a distinctly visible, agglutination reaction occurs. In a serum with no such RF factors there will be no agglutination and latex suspension will be smooth and uniform.

DIAGNOSTIC SIGNIFICANCE:

RFs are almost always absent in rheumatic fever; hence the test is useful in differentiating between rheumatic fever and rheumatoid arthritis.

Latex, positive test shows presence of rheumatoid factors in significant quantities, even though latex negative results do not rule out rheumatoid arthritis. The semiquantitative test will give concentration of these factors and requires an experts interpretation using other clinical findings.

Rheumatoid factors are present in serum from patients of rheumatoid arthritis, SLE, Reiter's syndrome, gout, psoriatic arthritis etc. Positive results may occur, sometimes in various pathological diseases including hepatitis, cirrhosis, lymphomas etc.

PRESENTATION:

	No. of vials/pack		
	25 Test / 50 Tests	100 Test	
1. Latex Reagent	1		} Provided as per pack size
2. Positive control serum	1		
3. Test Slides			
4. Mixing Sticks			
5. Plastic Droppers			
6. Glass Droppers			

REAGENT STORAGE AND STABILITY:

The ready to use Latex reagent is stable up to the expiry date printed on the label when stored at 2-8° C. **DO NOT FREEZE.**

SPECIMEN:

Fresh serum should be used. Serum sample is stable at - 20°C for 4 weeks. Plasma should not be used as fibrinogen may cause nonspecific agglutination of the latex particles. Do not use lipemic hemolysed or contaminated serum.

PROCEDURE:

A) QUALITATIVE TEST :

- 1) Allow reagent and sample to come at room temperature before use.

- 2) Mix latex reagent well before use.
- 3) Identify the circles on the slide.
- 4) Place approx. 25 µl. of each sample into the center of the circle marked for the sample on the slide.
- 5) Mix latex reagent well. Hold the dropper vertical and add equal amount of the latex suspension to each sample or controls. Spread over the circle using separate mixing sticks for each sample.
- 6) Rock the slide back and forth gently for two minutes or place the slide on an automated rotator at 100 rpm.

PROCEDURE AT A GLANCE

25 µl. of sample or control
+
25 µl. of RF Latex Reagent
Mix

Rock the slide; observe for agglutination up to 2 minutes

INTERPRETATION OF RESULTS:

Latex agglutination indicated that the RF (Rheumatoid factor) level is higher than 8 IU/ml. Sera showing positive results in screening test may subject to semi quantitative test.

B) SEMI-QUANTITATIVE TEST:

Prepare serial dilution 1:2, 1:4, 1:8, 1:16, 1:32 of the positive serum sample using physiological saline (0.9 %). Test the diluted sample using qualitative procedure and check for agglutination.

The approximate RF level in the sample can be calculated as follows.

$$RF \text{ (IU/ml.)} = (\text{Highest dilution showing positive reaction}) \times [\text{sensitivity of reagent (8 IU/ml.)}]$$

SENSITIVITY:

The reagent has a sensitivity of 8 IU/ml.

QUALITY CONTROL PROCEDURE:

The use of positive and negative controls are recommended along with serum sample.

PRECAUTIONS:

1. Drying of the test mixture at the periphery of the circle may lead to erroneous interpretation of results.
2. Read and interpret results exactly at 2 minutes.
3. As with all diagnostic methods, the final diagnosis should not be made using the result of a single test but should be correlated with other clinical findings.

REFERENCE:

1. Bendilla, K. L., and Mc.Duffie, F. C., Arthritis Rheum, 12 (1969) 74.
2. Muller, W., The serology.
3. Muller, W., The serology of Rheumatoid Arthritis, Berlin, Gottingen, Heidelberg (1962) 97.
4. Scott Det al : Systemic rheumatoid vasculitis: A clinical and laboratory study of 50 cases. Medicine 1981:60-288. Zvaifler NJ: Rheumatoid synovitis: An extravascular immune complex disease. Arthritis Rheum 1974; 17:297.
5. Moore T. Weiss T : Immunologic studies in juvenile arthritis. Bull Rheum Dis 1982; 32 : 25.

